BESCHEINIGUNG ◆ ATTESTATION ◆ 证明书 ◆ СВИДЕТЕЛЬСТВО ◆ CONSTANCIA ◆ ATTESTAZIONE

TÜV SÜD Türkiye

Conformance Certificate

Certificate No. : **Z-IS-TGK-IST-25-07-5010011896-25-IS-0466a-34-T-03-ENG**

User : **NORM Bağlantı ve Tesbit Elemanları San. Tic. Ltd. Şti.**
Hadımköy- Arnavutköy - İstanbul /TÜRKİYE

Description of Equipment : **Tensile Test Machine with Load Cell**

Model &Type & Capacity : **Test Machine 2022 50 kN**

Manufacturer : **MİCROANALİZ**

Serial No : 170900143

Sensivity : 0,001 kN

Calibration No / Date : 0759.NOR.01 - 27.02.2025

Calibrated By : Emis Kalibrasyon ve Ölçüm Hizm. San. Tic. Ltd. Şti.
accredited by TÜRKAK

Measuring Uncertainity : %0,21

Inspection Date & Place : 03.07.2025
Norm Bağlantı Ltd.Şti - Hadımköy - Arnavutköy - İstanbul / TÜRKİYE

Test Standard : Customer Request

Result :
It was observed that the tensile test machine mentioned above has been calibrated by yearly and found to be satisfactory for performing tensile test.
This certificate has been issued upon request of Norm Bağlantı ve Tesbit Elemanları San. Tic. Ltd. Şti.

Certificates with a three-year term of validity require annual surveillance audits.
The certificate can be renewed if required and on request.

This certificate is valid until July 02th , 2028

Registration no: 25-SER-03359-IS_TP

Istanbul, 04.07.2025

TÜV SÜD TÜRKİYE
Ayşegül ALKAN

TÜV SÜD Türkiye Industry Services